



**PRİZMANET MEDİKAL SANAYİ TİCARET  
İTHALAT İHRACAT LİMİTED ŞİRKETİ**

# **OUR CERTIFICATES**

**CE**

**ISO 9001:2015**

**ISO 45001:2018**

**ISO 14001:2015**

**ISO 22716:2008:GMP**

**ISO 13485:2016**

**EN 14683:2019+AC:2019 TIP IIR  
TEST REPORT**



# EU Declaration of Conformity

## AB Uygunluk Beyanı

No. DoC-25052021-1

| Ürün Adı / Name of Product                                       | : | Tıbbi Yüz Maskesi (Tek Kullanımlık) (Üç Katmanlı) / Yetişkin / Çocuk<br>Medical Face Mask (Disposable) (Three-layer) / Adult / Child                                                                                                                                                                                                                                                                            |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |
|------------------------------------------------------------------|---|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------|---|----------|-------------------------------------------|---|------|---------------------------|---|------|----------------------------------|---|-------|-------------------------------|---|---|
| GMDN Kodu / GMDN Code                                            | : | 35177 (Surgical Face Mask, Single Use)                                                                                                                                                                                                                                                                                                                                                                          |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |
| Branş Türü Kodu / UMDNS Code                                     | : | 1229 (Masks, Surgical)                                                                                                                                                                                                                                                                                                                                                                                          |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |
| Yasal Üretici (Adresi) /<br>Legal Manufacturer (Place of Issue)  | : | <b>PRİZMANET MEDİKAL SAN. TİC. İTH. İHR. LTD. ŞTİ.</b><br>Yahya Kemal Mah. Okul Cad. No:13/15,<br>Kağıthane / İstanbul, Turkey                                                                                                                                                                                                                                                                                  |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |
| Marka / Trademark                                                | : | <b>Medizer – Mora Mask</b>                                                                                                                                                                                                                                                                                                                                                                                      |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |
| Deney Raporu / Test Report                                       | : | <b>Ekoteks Laboratuvar ve Gözetim Hiz. A.Ş.</b>                                                                                                                                                                                                                                                                                                                                                                 |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |
| Deney Rapor No / Test Report No                                  | : | <b>AB-0583-T 21005288 02-21</b>                                                                                                                                                                                                                                                                                                                                                                                 |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |
| Ürün Standartları / Product<br>Standards                         | : | EN 14683:2019+AC:2019                                                                                                                                                                                                                                                                                                                                                                                           |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |
| Özellikler / Parameters                                          | : | <table><thead><tr><th>EU Standard Class</th><th>:</th><th>Type IIR</th></tr></thead><tbody><tr><td>Bacterial Filtration Efficiency (BFE) (%)</td><td>:</td><td>98,7</td></tr><tr><td>Breathing Resistance (Pa)</td><td>:</td><td>47,1</td></tr><tr><td>Splash Resistance Pressure (kPa)</td><td>:</td><td>&gt;21,3</td></tr><tr><td>Microbial Cleanliness (cfu/g)</td><td>:</td><td>9</td></tr></tbody></table> | EU Standard Class | : | Type IIR | Bacterial Filtration Efficiency (BFE) (%) | : | 98,7 | Breathing Resistance (Pa) | : | 47,1 | Splash Resistance Pressure (kPa) | : | >21,3 | Microbial Cleanliness (cfu/g) | : | 9 |
| EU Standard Class                                                | : | Type IIR                                                                                                                                                                                                                                                                                                                                                                                                        |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |
| Bacterial Filtration Efficiency (BFE) (%)                        | : | 98,7                                                                                                                                                                                                                                                                                                                                                                                                            |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |
| Breathing Resistance (Pa)                                        | : | 47,1                                                                                                                                                                                                                                                                                                                                                                                                            |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |
| Splash Resistance Pressure (kPa)                                 | : | >21,3                                                                                                                                                                                                                                                                                                                                                                                                           |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |
| Microbial Cleanliness (cfu/g)                                    | : | 9                                                                                                                                                                                                                                                                                                                                                                                                               |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |
| Ürün Sınıflandırması, Kurallar /<br>Device Classification, Rules | : | <b>Sınıf I</b> , steril olmayan, ölçme fonksiyonu olmayan ve<br>invaziv olmayan, Ek VIII uyarınca Kural 1 ( <b>93/42/AT – Ek VII</b> )<br><b>Class I</b> , non-sterile, non-measuring and non-invasive,<br>Rule 1 per Annex VIII ( <b>Ex 93/42/EEC – Annex VII</b> )                                                                                                                                            |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |
| Belge No / CE Certificate Number                                 | : | N/A                                                                                                                                                                                                                                                                                                                                                                                                             |                   |   |          |                                           |   |      |                           |   |      |                                  |   |       |                               |   |   |

Bu AB Uygunluk Beyanı yalnızca üretici sorumluluğunda düzenlenmiştir. Aşağıda imzası bulunan ben, işbu belgeyle, belirtilen tıbbi cihazların 5 Nisan 2017 tarihli Tıbbi cihazlarla ilgili Konsey ve Avrupa Parlamentosu (AB) 2017/745 (Eski 93/42/AT) sayılı tüzüğünün ilgili hükümlerini karşıladığını beyan ederim. Yukarıda belirtilen Direktife uygunluğu gösteren tüm destekleyici belgeler "PRİZMANET MEDİKAL SAN. TİC. İTH. İHR. LTD. ŞTİ." firması tarafından korunmaktadır. Bu beyan, imza tarihinden sonraki tüm cihazlar için geçerlidir.

*This EU declaration of conformity is issued only under the responsibility of the manufacturer. I, the undersigned, hereby declare that the specified medical device(s) meet(s) the applicable provisions of Regulation (EU) 2017/745 (Ex 93/42/EEC) of the European Parliament and of the Council of 5 April 2017 on medical devices. All supporting documentation that contains proof of compliance to the aforementioned Directive(s) is retained under the premises of "PRİZMANET MEDİKAL SAN. TİC. İTH. İHR. LTD. ŞTİ. This declaration applies to all devices from the signature date forward.*

**CE İşaretleme Tarihi / AB Uygunluk Beyanı Tarihi**  
**Start of CE / EU Declaration of Conformity Date**

25.05.2021

**Yetkili İmzası, Authorized Signature**

PRİZMANET MEDİKAL SAN. TİC. İTH. İHR. LTD. ŞTİ.

**AHMET BİLAL ERBEK, Manager**

Istanbul, 25-May-2021

**PRİZMANET MEDİKAL SANAYİ TİCARET  
İTHALAT İHRACAT LİMİTED ŞİRKETİ**

YAHYA KEMAL MAH. OKUL CAD. NO:13/15  
KAĞITHANE / İSTANBUL / TÜRKİYE

*Has been assessed and found to Comply with the Requirements of:*  
*Denetlenmiş ve aşağıdaki standardın gerekliliklerine uygunluğu görülmüştür:*

**ISO 9001:2015**

*The Quality Management System is applicable to:*  
*Kalite Yönetim Sistemi:*

**PRODUCTION AND SALES OF WORKING CLOTHES AND SHOES AND  
MEDICAL MATERIALS**

**İŞ ELBİSELERİ VE AYAKKABILARI VE MEDİKAL MALZEMELERİN  
ÜRETİMİ VE SATIŞI**

**Certificate Number: QMS-11646/01**  
**Belge Numarası: QMS-11646/01**

**Initial Certification Date: 03.11.2020**  
**İlk Belgelendirme Tarihi: 03.11.2020**

**Certification Period: 3 Years**  
**Belgelendirme Periyodu: 3 Yıl**

**Certificate Validity Date: 02.11.2021**  
**Belge Geçerlilik Tarihi: 02.11.2021**



*IQR Sertifikasyon Onayı*



**ACCREDITED®**  
Management Systems  
Certification Body  
**MSCB-135**

**IQR ULUSLARARASI BELGELENDİRME HİZMETLERİ LTD.ŞTİ.**

Beşevler Mah. Kocayunus Sk. Arslan Han İş Merkezi K:2 Nilüfer / BURSA

Tel: +90 224 266 00 16 Faks: +90 224 249 41 13 www.iqrcert.com e-posta:info@iqrcert.com





## PRİZMANET MEDİKAL SANAYİ TİCARET İTHALAT İHRACAT LİMİTED ŞİRKETİ

YAHYA KEMAL MAH. OKUL CAD. NO:13/15  
KAĞITHANE / İSTANBUL / TÜRKİYE

*Has been assessed and found to Comply with the Requirements of:  
Denetlenmiş ve aşağıdaki standardın gerekliliklerine uygunluğu görülmüştür:*

### ISO 45001:2018

*The Occupational Health and Safety Management System is applicable to:  
İş Sağlığı Ve Güvenliği Yönetim Sistemi:*

**PURCHASE, SALES, IMPORT, EXPORT OF WORK CLOTHES, SHOES AND  
MEDICAL SUPPLIES, FACE SHIELDS, FFP2 MASKS, DISPOSABLE MASKS,  
SABOT SLIPPERS, CHILD MASKS, HEALTHCARE PERSONNEL OUTFITS,  
SURGICAL CAPS AND SURGICAL GLOVES**

**İŞ ELBİSELERİ, AYAKKABILARI VE MEDİKAL MALZEMELERİN, YÜZ  
SİPERLİĞİ, FFP2 MASKE, TEK KULLANIMLIK MASKE, SABO TERLİK, ÇOCUK  
MASKESİ, SAĞLIK PERSONELİ KIYAFETİ, CERRAHİ BONE VE CERRAHİ  
ELDİVEN ALIMI, SATIMI, İTHALATI, İHRACATI**

Certificate Number: ISO-05044  
Belge Numarası: ISO-05044

Initial Certification Date: 15.02.2021  
İlk Belgelendirme Tarihi: 15.02.2021

Certification Period: 3 Years  
Belgelendirme Periyodu: 3 Yıl

Certificate Validity Date: 14.02.2022  
Belge Geçerlilik Tarihi: 14.02.2022



*IQR Sertifikasyon Onayı*

**IQR ULUSLARARASI BELGELENDİRME HİZMETLERİ LTD.ŞTİ.**

Beşevler Mah. Kocayunus Sk. Arslan Han İş Merkezi K:2 Nilüfer / BURSA  
Tel: +90 224 266 00 16 Faks: +90 224 249 41 13 www.iqrcert.com e-posta:info@iqrcert.com





### PRİZMANET MEDİKAL SANAYİ TİCARET İTHALAT İHRACAT LİMİTED ŞİRKETİ

YAHYA KEMAL MAH. OKUL CAD. NO:13/15  
KAĞITHANE / İSTANBUL / TÜRKİYE

*Has been assessed and found to Comply with the Requirements of:*  
*Denetlenmiş ve aşağıdaki standardın gerekliliklerine uygunluğu görülmüştür:*

## ISO 14001:2015

*The Environmental Management System is applicable to*  
*Çevre Yönetim Sistemi:*

**PURCHASE, SALES, IMPORT, EXPORT OF WORK CLOTHES, SHOES AND  
MEDICAL SUPPLIES, FACE SHIELDS, FFP2 MASKS, DISPOSABLE MASKS,  
SABOT SLIPPERS, CHILD MASKS, HEALTHCARE PERSONNEL OUTFITS,  
SURGICAL CAPS AND SURGICAL GLOVES**

**İŞ ELBİSELERİ, AYAKKABILARI VE MEDİKAL MALZEMELERİN, YÜZ  
SİPERLİĞİ, FFP2 MASKE, TEK KULLANIMLIK MASKE, SABO TERLİK, ÇOCUK  
MASKESİ, SAĞLIK PERSONELİ KIYAFETİ, CERRAHİ BONE VE CERRAHİ  
ELDİVEN ALIMI, SATIMI, İTHALATI, İHRACATI**

Certificate Number: EMS-12879  
Belge Numarası: EMS-12879

Initial Certification Date: 15.02.2021  
İlk Belgelendirme Tarihi: 15.02.2021

Certification Period: 3 Years  
Belgelendirme Periyodu: 3 Yıl

Certificate Validity Date: 14.02.2022  
Belge Geçerlilik Tarihi: 14.02.2022

IQR Sertifikasyon Onayı



ACCREDITED<sup>®</sup>  
Management Systems  
Certification Body  
MSCB-135

**IQR ULUSLARARASI BELGELENDİRME HİZMETLERİ LTD.ŞTİ.**

Beşevler Mah. Kocayunus Sk. Arslan Han İş Merkezi K:2 Nilüfer / BURSA  
Tel: +90 224 266 00 16 Faks: +90 224 249 41 13 www.iqrcert.com e-posta:info@iqrcert.com





# Certificate



## PRİZMANET MEDİKAL SANAYİ TİCARET İTHALAT İHRACAT LİMİTED ŞİRKETİ

YAHYA KEMAL MAH. OKUL CAD. NO:13/15  
KAĞITHANE / İSTANBUL / TÜRKİYE

*Has been assessed and found to Comply with the Requirements of:  
Denetlenmiş ve aşağıdaki standartın gerekliliklerine uygunluğu görülmüştür:*

**ISO 22716:2008:GMP**  
**GOOD MANUFACTURING PRACTICES**

*The GMP System is applicable to:  
İyi Üretim Uygulamaları:*

**PURCHASE, SALES, IMPORT, EXPORT OF WORK CLOTHES, SHOES AND  
MEDICAL SUPPLIES, FACE SHIELDS, FFP2 MASKS, DISPOSABLE MASKS,  
SABOT SLIPPERS, CHILD MASKS, HEALTHCARE PERSONNEL OUTFITS,  
SURGICAL CAPS AND SURGICAL GLOVES**

**İŞ ELBİSELERİ, AYAKKABILARI VE MEDİKAL MALZEMELERİN, YÜZ  
SİPERLİĞİ, FFP2 MASKE, TEK KULLANIMLIK MASKE, SABO TERLİK, ÇOCUK  
MASKESİ, SAĞLIK PERSONELİ KIYAFETİ, CERRAHİ BONE VE CERRAHİ  
ELDİVEN ALIMI, SATIMI, İTHALATI, İHRACATI**

Certificate No: ISO/23181  
Sertifika Numarası: ISO/23181

Date of Issue: 15.02.2021  
Yayınlanma Tarihi: 15.02.2021

1st Surveillance Audit : February 2022  
1. Gözetim Denetim Tarihi : Şubat 2022  
2nd Surveillance Audit : February 2023  
2. Gözetim Denetim Tarihi : Şubat 2023

Date of Certification Audit: 08.02.2021  
Denetim Tarihi: 08.02.2021

Date of Expiry: 14.02.2022  
Son Geçerlilik Tarihi: 14.02.2022



Qcs International Certifications Services Pvt. Ltd  
Certificate of validity information E-mail: info@qcsert.com  
mail to confirm



**PRİZMANET MEDİKAL SANAYİ TİCARET  
İTHALAT İHRACAT LİMİTED ŞİRKETİ**

# ISO 13485:2016

**YETİŞKİN TIBBİ MASKESİ, ÇOCUK TIBBİ MASKESİ, BONE, SİPERLİK, TULUM  
ÜRETİMİ VE SATIŞI**

|                                                 |                        |
|-------------------------------------------------|------------------------|
| <b>Date of initial registration</b>             | <b>08 October 2020</b> |
| <b>Date of this Certificate</b>                 | <b>08 October 2020</b> |
| <b>Surveillance audit on or before</b>          | <b>07 October 2021</b> |
| <b>Recertification Due / Certificate expiry</b> | <b>07 October 2023</b> |



*Emmanuel*  
**Emmanuel ADEMOSU**  
**Director**





|                  |
|------------------|
| AB-0583-T        |
| 21005288<br>-ing |
| 02-21            |

**Customer name:** PRİZMA NET MEDİKAL SAN. TİC. İHR. İTH. LTD. ŞTİ.

**Address:** YAHYA KEMAL MAH. OKUL CAD. NO:13/15 K.HANE/İST.

**Buyer name:**

**Contact Person:** -

**Order No:** -

**Article No:** MEDIZER MB001

**Name and identity of test item:** Blue non woven mask (Claimed to be;70PIECES-WHITE)

**The date of receipt of test item:** 10.02.2021

**Re-submitted/re-confirmation date:** -

**Date of test:** 10.02.2021-15.02.2021

**Remarks:** -


**Sampling:** The results given in this report belong to the received sample by vendor.

**End-Use:** -

**Care Label:** -

**Number of pages of the report:** 5

The Turkish Accreditation Agency (TURKAK) is signatory to the multilateral agreements of the European co-operation for the Accreditation (EA) and of the International Laboratory Accreditation (ILAC) for the Mutual recognition of test reports. EKOTEKS LABORATUVAR ve GÖZETİM HİZMETLERİ A.Ş. accredited by TÜRKAK under registration number [AB-0583-T] for ISO 17025:2017 as test laboratory. The test and/or measurement results, the uncertainties (if applicable) with confidence probability and test methods are given on the following pages which are part of this report.



**Date**  
15.02.2021

**Customer Representative**  
Servin YURTSEVEN

**Head of Testing Laboratory**  
Sevim A. RAZAK  
15.02.2021

This report shall not be reproduced other than in full except with the permission of the laboratory.  
Testing reports without signature and seal are not valid.



**EKOTEKS LABORATUVAR ve GÖZETİM  
HİZMETLERİ A.Ş.**

AB-0583-T

21005288  
-ing

02-21

| REQUIRED TESTS                                                                                                                        | RESULT | COMMENTS |
|---------------------------------------------------------------------------------------------------------------------------------------|--------|----------|
| MICROBIOLOGICAL TESTS                                                                                                                 |        |          |
| Bacterial Filtration Efficiency-BFE                                                                                                   | P      | TYPE IIR |
| Microbial Cleanliness(Bioburden)                                                                                                      | P      |          |
| PHYSICAL PROPERTIES                                                                                                                   |        |          |
| Breathability(Differential Pressure)                                                                                                  | P      |          |
| Blood Splash Resistance                                                                                                               | P      |          |
| P: Pass<br>F: Fail<br>R: Refer to retailer technologist<br><br>Test results evaluated according to EN 14683:2019+AC:2019 limit values |        |          |

**REMARK:** Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified.If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95 %. The declaration of conformity was given in accordance with the Simple Acceptance Decision Rule. Tests marked (\*) in this report are not included in the accreditation schedule.



This report shall not be reproduced other than in full except with the permission of the laboratory.  
*Testing reports without signature and seal are not valid.*



**EKOTEKS LABORATUVAR ve GÖZETİM  
HİZMETLERİ A.Ş.**

AB-0583-T

21005288  
-ing

02-21

## TEST RESULT

### Medical face masks - Requirements and test methods

#### EN 14683:2019+AC:2019 (TS EN 14683+AC:2019)

#### BACTERIAL FILTRATION EFFICIENCY (BFE)

**Test Metodu:** EN 14683:2019+AC :2019 (TS EN 14683+AC:2019)

A specimen of the mask material is clamped between an impactor and an aerosol chamber. An aerosol of *Staphylococcus aureus* is introduced into the aerosol chamber and drawn through the mask material and the impactor under vacuum. The bacterial filtration efficiency of the mask is given by the number of colony forming units passing through the medical face mask material expressed as a percentage of the number of colony forming units present in the challenge aerosol.

|                                                           |                                                       |
|-----------------------------------------------------------|-------------------------------------------------------|
| Test Flow Rate                                            | 28,3 L/min                                            |
| Total Test Flow Time                                      | 2 minute                                              |
| Sample Sizes                                              | 5 pieces mask                                         |
| Test Alanı                                                | 4.9 cm <sup>2</sup> (5 replicas)                      |
| Test Condition                                            | (21 ± 5) °C and (85 ± 5) % relative humidity, 4 hours |
| Test Microorganism                                        | <i>Staphylococcus aureus</i> ATCC 6538                |
| Bacterial concentration (cfu/ ml )                        | 5x10 <sup>5</sup> cfu/ ml                             |
| incubation conditions                                     | 24 hour, 35°C ± 2°C                                   |
| Positive control sample average of number of Bacteria (C) | 3 x10 <sup>3</sup> cfu/ ml                            |
| Mean particle size (MPS)                                  | 3.0 µm                                                |

| RESULTS               |                                                |                                            |                               |
|-----------------------|------------------------------------------------|--------------------------------------------|-------------------------------|
| Number of Test Sample | Test Sample (T)<br>Number of Bacteria<br>(cfu) | Bacterial Filtration<br>Efficiency ( % B ) | Requirement<br>BFE (%)        |
| 1                     | 39                                             | %98.7                                      | Type I ≥95<br><br>Type II ≥98 |
| 2                     | 42                                             | %98.6                                      |                               |
| 3                     | 45                                             | %98.5                                      |                               |
| 4                     | 46                                             | %98.5                                      |                               |
| 5                     | 47                                             | %98.4                                      |                               |

cfu: Colony-forming unit

$$B = (C - T) / C \times 100$$

%B: Bacterial Filtration Efficiency

C: is the mean of the total plate counts for the two positive control runs

T: is the total plate count for the test specimen



**EKOTEKS LABORATUVAR ve GÖZETİM  
HİZMETLERİ A.Ş.**

AB-0583-T

21005288  
-ing

02-21

**TEST RESULT**

**BREATHABILITY (Differential Pressure)**

**Test Metodu:** EN 14683:2019+AC :2019 (TS EN 14683+AC:2019)

Test Condition (21 ± 5) °C ve (85 ± 5) % relative humidity, 4 hrs

Test area is 25 mm in diameter , 5 different sample was taken

Adjusted airflow is 8 l/min. The differential pressure is read directly using a differential pressure manometer .

| SAMPLE         | DIFFERENTIAL<br>PRESSURE RESULT | REQUIREMENT             |
|----------------|---------------------------------|-------------------------|
| 1              | 39.7 Pa/cm <sup>2</sup>         | < 60 Pa/cm <sup>2</sup> |
| 2              | 47.3 Pa/cm <sup>2</sup>         |                         |
| 3              | 46.3 Pa/cm <sup>2</sup>         |                         |
| 4              | 49.0 Pa/cm <sup>2</sup>         |                         |
| 5              | 53.3 Pa/cm <sup>2</sup>         |                         |
| Average Result | 47.1 Pa/cm <sup>2</sup>         |                         |

**MICROBIAL CLEANLINESS (Bioburden)**

**Test Metod:** EN 14683:2019+AC :2019 (TS EN 14683+AC:2019)  
EN ISO 11737-1:2018 /TS EN ISO 11737-1 :2018

5 sample were taken. The sample is weighted and put in extraction liquid after shaking well (250 rpm, 5 min), inoculated on the suitable agar.

The plates are incubated for 3 days at 30 ± 1 °C for 72 hours, and 7 days at (20 to 25) °C for TSA and SDA plates respectively. Total microorganisms counts are calculated.

|                               | RESULTS | REQUIREMENT |
|-------------------------------|---------|-------------|
| Microbial cleanliness (cfu/g) | 9 cfu/g | ≤30 cfu/g   |

\*cfu= Colony forming unit.



**EKOTEKS LABORATUVAR ve GÖZETİM  
HİZMETLERİ A.Ş.**

AB-0583-T

21005288  
-ing

02-21

## TEST RESULT

### BLOOD SPLASH RESISTANCE

**Test Metod:** EN 14683:2019+AC :2019 (Clause 5.2.4) the resistance of the medical face mask to penetration  
ISO 22609 :2004 Clothing for protection against infectious agents — Medical face masks — Test method for resistance against penetration by synthetic blood (fixed volume, horizontally projected)  
**Test Condition** (21 ± 5) °C ve (85 ± 5) % relative humidity, 4 hrs  
32 different sample was taken

|                       | <b><u>SPLASH RESISTANCE<br/>PRESSURE (kPa)</u></b> | <b><u>RESULTS</u></b> | <b><u>REQUIREMENT</u></b>       |
|-----------------------|----------------------------------------------------|-----------------------|---------------------------------|
| 1                     | >21.3 kPa                                          | PASS                  | <b>≥16 kPa</b><br>Type IIR mask |
| 2                     | >21.3 kPa                                          | PASS                  |                                 |
| 3                     | >21.3 kPa                                          | PASS                  |                                 |
| 4                     | >21.3 kPa                                          | PASS                  |                                 |
| 5                     | >21.3 kPa                                          | PASS                  |                                 |
| 6                     | >21.3 kPa                                          | PASS                  |                                 |
| 7                     | >21.3 kPa                                          | PASS                  |                                 |
| 8                     | >21.3 kPa                                          | PASS                  |                                 |
| 9                     | >21.3 kPa                                          | PASS                  |                                 |
| 10                    | >21.3 kPa                                          | PASS                  |                                 |
| 11                    | >21.3 kPa                                          | PASS                  |                                 |
| 12                    | >21.3 kPa                                          | PASS                  |                                 |
| 13                    | >21.3 kPa                                          | PASS                  |                                 |
| 14                    | >21.3 kPa                                          | PASS                  |                                 |
| 15                    | >21.3 kPa                                          | PASS                  |                                 |
| 16                    | >21.3 kPa                                          | PASS                  |                                 |
| 17                    | >21.3 kPa                                          | PASS                  |                                 |
| 18                    | >21.3 kPa                                          | PASS                  |                                 |
| 19                    | >21.3 kPa                                          | PASS                  |                                 |
| 20                    | >21.3 kPa                                          | PASS                  |                                 |
| 21                    | >21.3 kPa                                          | PASS                  |                                 |
| 22                    | >21.3 kPa                                          | PASS                  |                                 |
| 23                    | >21.3 kPa                                          | PASS                  |                                 |
| 24                    | >21.3 kPa                                          | PASS                  |                                 |
| 25                    | >21.3 kPa                                          | PASS                  |                                 |
| 26                    | >21.3 kPa                                          | PASS                  |                                 |
| 27                    | >21.3 kPa                                          | PASS                  |                                 |
| 28                    | >21.3 kPa                                          | PASS                  |                                 |
| 29                    | >21.3 kPa                                          | PASS                  |                                 |
| 30                    | >21.3 kPa                                          | PASS                  |                                 |
| 31                    | >21.3 kPa                                          | PASS                  |                                 |
| 32                    | >21.3 kPa                                          | PASS                  |                                 |
| <b>Average Result</b> | >21.3 kPa                                          | PASS                  |                                 |

Gen.f136-2/03